




Project Number: na		Tracking Code: TC0332--0246	
Requested by: Mark Shireman		Date: 8/7/2003	Product Rev: N/A
Part #: MMS-140-02-L-DV		Lot #: N/A	Tech: Troy Cook Eng: John Tozier
Part description: Box Contact Socket Strip			Qty to test: 2
Test Start: 08/07/2003	Test Completed: 8/7/2003		



Representative Sample, MMS-105 depicted

CCC Report

PART DESCRIPTION

MMS-140-02-L-DV

Mated with TMM-140-01-S-D-SM

CERTIFICATION

All instruments and measuring equipment were calibrated to National Institute for Standards and Technology (NIST) traceable standards according to ISO 10012-1 and ANSI/NCSL 2540-1, as applicable.

All contents contained herein are the property of Samtec. No portion of this report, in part or in full shall be reproduced without prior written approval of Samtec.

SCOPE

To perform the following tests: CCC

APPLICABLE DOCUMENTS

Standards: EIA Publication 364

TEST SAMPLES AND PREPARATION

- 1) All materials were manufactured in accordance with the applicable product specification.
- 2) All test samples were identified and encoded to maintain traceability throughout the test sequences.
- 3) After soldering, the parts to be used for LLCR and DWV/IR testing were cleaned according to TLWI-0001.
- 4) Either an automated cleaning procedure or an ultrasonic cleaning procedure may be used.
- 5) The automated procedure is used with aqueous compatible soldering materials.
- 6) The ultrasonic procedure can be used with either aqueous or non-aqueous soldering components and follows:
 - a) Sample test boards are to be ultrasonically cleaned after test lead attachment, preparation and/or soldering.
 - b) Sample test boards are immersed into Branson 3510 cleaner containing Kyzen Ionox HC1 (or equivalent) with the following conditions:
 - i) Temperature: -----55° C +/- 5° C
 - ii) Frequency:-----40 KHz
 - iii) Immersion Time: -----5 to 10 Minutes
 - c) Sample test boards are removed and placed into the Branson 3510 cleaner containing deionized water with the following conditions:
 - i) Temperature: -----55° C +/- 5° C
 - ii) Frequency:-----40 KHz
 - iii) Immersion Time: -----5 to 10 Minutes
 - d) Sample test boards are removed and placed in a beaker positioned on a hot plate with a magnetic stirrer containing deionized water warmed to 55° C +/- 5° C for 1/2 to 1 minute.
 - e) Upon removal, the sample boards are rinsed for 1/2 to 1 minute at room temperature with free flowing deionized water.
 - f) After the final rinse, the sample test boards are dried in an air-circulating oven for 10 to 15 minutes at 50° C +/- 5° C.
 - g) Sample test boards are then allowed to set and recover to room ambient condition prior to testing.
- 7) Parts not intended for testing LLCR and DWV/IR are visually inspected and cleaned if necessary.
- 8) Any additional preparation will be noted in the individual test sequences.
- 9) Solder Information: Kester #66/331, 63/37 with organic core.
- 10) Internal Test PCBs used: Product not mounted to PCB. Interconnections made by hand preparation.

ATTRIBUTE DEFINITION

The following is a brief, simplified description of attributes.

TEMPERATURE RISE (Current Carrying Capacity, CCC):

- 1) EIA-364-70, *Temperature Rise versus Current Test Procedure for Electrical Connectors and Sockets*.
- 2) When current passes through a contact, the temperature of the contact increases as a result of I^2R (resistive) heating.
- 3) The number of contacts being investigated plays a significant part in power dissipation and therefore temperature rise.
- 4) The size of the temperature probe can affect the measured temperature.
- 5) Copper traces on PC boards will contribute to temperature rise:
 - a) Self heating (resistive)
 - b) Reduction in heat sink capacity affecting the heated contacts
- 6) A de-rating curve, usually 20%, is calculated.
- 7) Calculated de-rated currents at three temperature points are reported:
 - a) Ambient
 - b) 60° C
 - c) 80° C
 - d) 95° C
- 8) Typically, neighboring contacts (in close proximity to maximize heat build up) are energized.
- 9) The thermocouple (or temperature measuring probe) will be positioned at a location to sense the MAXIMUM temperature in the vicinity of the heat generation area.
- 10) A computer program, *TR 803.exe*, ensures accurate stability for data acquisition.
- 11) Hook-up wire cross section is larger than the cross section of any connector leads/PC board traces, jumpers, etc.
- 12) Hook-up wire length is longer than the minimum specified in the referencing standard.

Tracking Code: TC0332--0246

Part #: MMS-140-02-L-DV

Part description: Box Contact Socket Strip


RESULTS

Temperature Rise, CCC at 20% de-rated

- At 95°C, relative to 125°C-----3.1 A with 6 adjacent contacts powered

DATA SUMMARY**TEMPERATURE RISE (Current Carrying Capacity, CCC):**

- 1) High quality thermocouples whose temperature slopes track one another were used for temperature monitoring.
- 2) The thermocouples were placed at a location to sense the MAXIMUM temperature generated during testing.
- 3) Temperature readings recorded are those for which three successive readings, 15 minutes apart, differ less than 1° C (computer controlled data acquisition).
- 4) Adjacent contacts were powered:
 - a) Linear configuration with six adjacent conductors/contacts powered



EQUIPMENT AND CALIBRATION SCHEDULES**Equipment #:** THL-01**Description:** Temperature/Humidity Chart Recorder**Manufacturer:** Dickson**Model:** THDX**Serial #:** 9316255**Accuracy:** Temp: +/- 1C; Humidity: +/-2% RH (0 - 60%) +/- 3% RH (61 - 95%).

... Last Cal: 6/15/2004, Next Cal: 6/15/2005

Equipment #: PS-01**Description:** System Power Supply**Manufacturer:** Hewlett Packard**Model:** HP 6033A**Serial #:** (HP) 3329A-07330**Accuracy:** See Manual 10/16/02- Had a fuse replaced and equipment was re-calibrated.

... Last Cal: 6/12/03, Next Cal: 6/12/04

Equipment #: MO-02**Description:** Multimeter /Data Acquisition System**Manufacturer:** Keithley**Model:** 2700**Serial #:** 0780546**Accuracy:** See Manual

... Last Cal: 6/12/03, Next Cal: 6/12/04

Equipment #: MO-04**Description:** Multimeter /Data Acquisition System**Manufacturer:** Keithley**Model:** 2700**Serial #:** 0798688**Accuracy:** See Manual

... Last Cal: 6/12/03, Next Cal: 6/12/04

Equipment #: TC090601-103/105**Description:** IC Thermocouple-103/105**Manufacturer:** Samtec**Serial #:** TC090601-103/105**Accuracy:** +/- 1 degree C